



Infusion System P50 Update Kit (1021421)

P50 INFUSION SYSTEM UPDATE KIT (1021421)

This update kit will enable you to improve your 3B Scientific® I.V. Injection Arm (P50).

Improvement of the infusion system:

- The infusion system of the I.V. injection arm is now easier to install.
- The infusion system is now a closed system reducing the risk of leakage during use.

Improvement of the artificial blood:

- The blood color is more realistic.
- Any stains caused by the new artificial blood can be washed out at 40°C.

Increase of efficiency:

- The new bottle of 250ml allows 10 applications, 6 more than the previous bottle.

▶ CONTENTS OF THE P50 INFUSION SYSTEM UPDATE KIT

- 1 stand
- 1 infusion bottle with a tube locking clip and a Y-tube connector
- 1 transparent tube connector
- 3 infusion tube system (as replacement parts)
- 1 bottle of 3B Scientific® artificial blood concentrate (250ml)



Old infusion system



› INSTRUCTION FOR USE

How to install the update kit easily



1. Remove and dispose of the old infusion system as seen on the previous picture
2. Plug the transparent tube connector on the shorter tube
3. Install the locking clip on the longer tube
4. Plug the Y-connector on the open tube at the bottom of the infusion bottle
5. Remove and dispose of the old metal stand and replace it by the new one.

Filling of the infusion bottle with the blood


1. Make sure that the locking clip of the ventilation tube of the infusion bottle is locked
2. Fill the infusion bottle up to the red line with the artificial blood concentrate
3. Then fill the infusion bottle up to the blue line with water
4. Close the bottle and shake gently to obtain a homogeneous solution.



P50 INFUSION SYSTEM UPDATE KIT (1021421)

Installation of the infusion system

1. Fasten the hanging hook at the middle of the stand
2. Hang the injection arm onto the hanging hook using the arm ring screw
3. Connect the shorter injection tube to the infusion bottle
4. Make sure that the locking clip of the longer tube is locked
5. Attach the infusion bottle upside down to the retaining clip by pushing the sealing cap of the bottle sideways into the grey retaining clip of the stand
6. Insert the longer tube in the plastic cup and allow the blood to flow into the cup for a short time to remove any bubbles
7. After checking that all the connections are correctly placed and that the longer tube is in the plastic cup, open the locking clip on the ventilation tube of the infusion bottle and the locking clip of the longer tube
8. The system will then be filled up and blood will start to flow out of the longer tube in the plastic cup. Let the blood flow out until you see no more bubbles coming out from the longer tube
9. Close the locking clip of the longer tube and connect it to one branch of the Y-connector on the infusion bottle
10. You can now start puncturing the peripheral veins.



After using of the Injection arm



To remove the artificial blood after use please follow the following steps:

1. Make sure that the locking clip of the ventilation tube of the infusion bottle is closed
2. Make sure that the locking clip of the longer tube is closed
3. Disconnect the longer tube from the Y-connector
4. Remove the infusion bottle from the stand retaining clip and turn the bottle upright
5. Disconnect the shorter tube from the bottle connector by gently unscrewing the connector
6. Put the ends of the tubes in a container or bucket in order to remove the remaining blood in the tubing system

If you are not using the injection arm for a prolonged period of time, it is recommended to rinse the complete tubing system so that no dye residues can settle in the system.

Please fill water into the infusion bottle and repeat the all installation steps in order to clean the system.

› CHANGING THE SKIN AND THE TUBING SYSTEM

How to change the tubing system

1. To change the tubing system you first need to remove the skin. Please remove first the transparent tube connector from the shorter tube and remove the locking clip from the longer tube.
2. Unscrew the ring screw and open the two parts of arm sealing caps



P50 INFUSION SYSTEM UPDATE KIT (1021421)



3. Starting at the top of the arm, roll the skin off the arm core
4. Unfasten the Velcro strips and remove the tubing system from the arm core
5. Fit the new tubing system into the provided grooves and refasten it using the Velcro strips.

How to put back the skin

1. The new skin can be pulled over the hand and arm core by putting some talcum powder onto the skin and distributing the powder by moving the skin back-and-forth.
2. Pull the skin over the hand core and arm core, put the tube back through the location hole and the cover plates. The rim of the skin should be closed between the two cover plates. Fasten the ring screw and replace the sealing caps onto the tubes.
3. Please don't forget to plug back the transparent tube connector on the shorter arm tube and to reinstall the locking clip on the longer tube.

▶ PLEASE NOTE THE FOLLOWING

- The skin and the tubing system of the injection arm are subject to wear and tear caused by usage. We supply the skin and the tubing system as a replacement parts.
- As the injection arm is an anatomical model, several – non-puncturable – veins are portrayed in the skin in addition to the puncturable veins.
- The artificial blood may cause stains. But don't worry, stains caused by the 3B Scientific® artificial blood can be washed out at 40°C.
- Indwelling vein cannulas may damage the material of the skin and of the tubing system. Please use injection cannulas of the recommended size (20 or 21 gauge) only and change the skin or the tubing system when damages of the skin become visible.

› PRODUCT NUMBERS FOR REPLACEMENT PARTS

Product Reference	Intern Reference	Description	Composition
1021419	(XP50/1-002)	Puncture accessories	1 Tubing System 1 Tube Connector 1 Tube Locking Clip 1 Disposable Syringe 2 Injection Cannulas 20 and 21 Gauge 1 Plastic Cup
1021422	(XP50/1-003)	Arm and hand core	1 Arm Core 1 Hand Core
1021423	(XP50/1-004)	Infusion bottle	1 Infusion Bottle
1021425	(XP50/1-005)	Stand	1 Stand
1021426	(XP50/1-006)	Replacement skin without core	13B SKINlike™ Silicone Arm Skin 1 Container of Talcum Powder
1021427	(XP50/1-007)	3 Tubing Systems	3 Tubing System 3 Tube Connector 3 Tube Locking Clip
1021251	(XP110)	Artificial Blood Concentrate	250 ml Bottle

› CONTACTING 3B SCIENTIFIC

When contacting technical support, please have your simulation product and manual available.



UK 3B Scientific Limited
Suite 1, Formal House, Oldmixon Crescent
Weston-super-Mare • BS24 9AY
Phone: 01934 425333 • Fax: 01934 425334
3bscientific.com • uk3bs@3bscientific.com